



# Microbatch Screening for Oryx Nano

## Step-by-Step Instructions

1. Switch to the 'Basic' Tab
2. Choose a Destination Plate type (there is only one)
3. Specify the Protein volume for the drops
4. Specify the Screen volume for the drops
5. The calculated % values are shown to the right of the drop
6. Enter a Protein name or description for the experiment if desired. This can be printed and will be logged
7. Check items on the 'Advanced' tab if desired
8. (Optional) **a.** Drag out a block on the Source Plate in the platelader view to dispense all or part of a plate. **b.** Click on the Destination Plate to locate the top left-hand corner of the block if it is smaller than the plate.
9. Click on the 'Dispense' menu item or button to execute the experiment.



WaspRun



Platelader View